





Configuration proposal Anesthesia (AN5)


Side view right





Front view



Side view left



Back



Configuration description:
 MediBoom®XXL 800/1000 with Anesthesia Machine Lifting
 1 x Shelf 520 mm, 2 x Drawer, 2 x IV pole w. extension

Brake system: Electromagnetic

1. Action space*: >4,400 mm
2. Weights and Pay Loads
 Remaining Net Pay Load: 237 kg
 Max. pendant pay load: 340 kg
 Vertical force as configured: 5684 N
3. Remaining Diameter: 41 %


					Number of service components						
O ₂	Air 4/5	N ₂ O	Vac	CO ₂	N ₂	AGSS	Air 8/ Air- motor	Electrical sockets	Earth bonding sockets	Double RJ45	Data prep.
2	2	2	2			1		10	6	2	3
Phoenix P3 British Standard (ISO)								Wands- worth			

* reference: centre of main bearing to centre of shelf




Configuration proposal Anesthesia (AN5)


Side view right



Front view



Side view left



Back

