





Configuration proposal General Surgery (CH4)



Configuration description:
 MediLift 1000/1000 with Service Head 600
 3 x Shelf 520 mm, 1 x Drawer

1. Brake system: Electromagnetic
2. Action space*: >4,800 mm
3. Weights and Pay Loads
 Remaining Net Pay Load: 94 kg
 Max. pendant pay load: 150 kg
 Vertical force as configured: 3166 N
4. Remaining Diameter: 58 %

Number of service components											
O ₂	Air 4/5	N ₂ O	Vac	CO ₂	N ₂	AGSS	Air 8/ Air- motor	Electrical sockets	Earth bonding sockets	Double RJ45	Data prep.
1	3		3				1	8	8	2	2
DKD AFNOR (Norme francaise)								Legrand			

* reference: centre of main bearing to centre of shelf




Configuration proposal General Surgery (CH4)


Side view right



Front view



Side view left



Back

